Validation of the da Vinci Surgical Skill Simulator across three surgical disciplines: A pilot study

Tarek Alzahrani, MD;^{*} Richard Haddad, MD;^{*} Abdullah Alkhayal, MD;^{*} Josée Delisle, MSc;^{*} Laura Drudi, MD;[†] Walter Gotlieb, MD;[†] Shannon Fraser, MD;[§] Simon Bergman, MD,[§] Frank Bladou, MD;^{*} Sero Andonian, MD, MSc, FRCSC;^{**} Maurice Anidjar, MD, PhD^{**}

*Department of Urology, McGill University, Montreal, QC; †Department of Gynecologic and Surgical Oncology, McGill University, Montreal, QC; §Department of General Surgery, McGill University, Montreal, QC; *Co-senior authors

Cite as: *Can Urol Assoc J* 2013;7(7-8):e520-9. http://dx.doi.org/10.5489/cuaj.419 Published online July 2, 2013.

Abstract

Objective: In this paper, we evaluate face, content and construct validity of the da Vinci Surgical Skills Simulator (dVSSS) across 3 surgical disciplines.

Methods: In total, 48 participants from urology, gynecology and general surgery participated in the study as novices (0 robotic cases performed), intermediates (1-74) or experts (≥75). Each participant completed 9 tasks (Peg board level 2, match board level 2, needle targeting, ring and rail level 2, dots and needles level 1, suture sponge level 2, energy dissection level 1, ring walk level 3 and tubes). The Mimic Technologies software scored each task from 0 (worst) to 100 (best) using several predetermined metrics. Face and content validity were evaluated by a questionnaire administered after task completion. Wilcoxon test was used to perform pair wise comparisons.

Results: The expert group comprised of 6 attending surgeons. The intermediate group included 4 attending surgeons, 3 fellows and 5 residents. The novices included 1 attending surgeon, 1 fellow, 13 residents, 13 medical students and 2 research assistants. The median number of robotic cases performed by experts and intermediates were 250 and 9, respectively. The median overall realistic score (face validity) was 8/10. Experts rated the usefulness of the simulator as a training tool for residents (content validity) as 8.5/10. For construct validity, experts outperformed novices in all 9 tasks (p < 0.05). Intermediates outperformed novices in 7 of 9 tasks (p < 0.05); there were no significant differences in the energy dissection and ring walk tasks. Finally, experts scored significantly better than intermediates in only 3 of 9 tasks (matchboard, dots and needles and energy dissection) (p < 0.05).

Conclusions: This study confirms the face, content and construct validities of the dVSSS across urology, gynecology and general surgery. Larger sample size and more complex tasks are needed to further differentiate intermediates from experts.

Introduction

The American College of Surgeons has recommended simulation-based training for surgical trainees to increase proficiency and patient safety in light of recent restrictions to trainee work hours and increased concerns over patient safety.1 The first validated low-fidelity objective training and assessment tool for basic laparoscopic skills is the McGill Inanimate System for Training and Evaluation of Laparoscopic Skills (MISTELS), which has been adapted by the Society of American Gastrointestinal and Endoscopic Surgeons (SAGES) for the Fundamentals of Laparoscopic Surgery (FLS) program.² Currently, FLS certification is required for candidates applying for certification from the American Board of Surgery. Furthermore, construct validity of MISTELS has been established for urology trainees and attending urologists.³ Recently, the American Urological Association's Committee on Laparoscopy, Robotic and New Technology modified the original MISTELS/FLS program into a more urocentric curriculum, the Basic Laparoscopic Urologic Surgery (BLUS) skills curriculum, by replacing the endoloop and extra-corporeal knot tying exercises with a novel clip-applying task.⁴ The construct validity of BLUS has been recently demonstrated for the assessment of basic laparoscopic skills of urologists.⁴

With the introduction of the da Vinci (Intuitive Surgical, Inc., Sunnyvale, CA) robotic-assisted laparoscopic surgery, the need for training and assessment of more complex robotic assisted-laparoscopic skills has increased. In one study comparing novice robotic surgeons performing the intracorporeal suturing exercise of the FLS program using standard laparoscopy versus da Vinci robotic assistance, participants found standard laparoscopy to be more physically demanding and preferred the robotic platform (62 vs. 38%, p < 0.01) despite achieving significantly higher suturing scores with standard laparoscopy when compared with robotic assistance.⁵ Therefore, there is increasing demand from both patients and

surgeons to incorporate this technology. The Ontario Health Technology Assessment found that despite high startup and maintenance cost, robotic-assisted laparoscopic hysterectomies and prostatectomies were associated with lower blood loss and hospitalization length.⁶ However, robotic-assisted laparoscopic surgery requires a unique set of skills that are not measured on the MISTELS, FLS or BLUS curricula.

Currently, there are three commercially available robotic platform simulators. The face and content validity of the Robotic Surgical Simulator (RoSS) (Simulated Surgical Systems; Williamsville, NY) have been demonstrated.^{7,8} The original software developed by Mimic Technologies for the second simulator, namely the da Vinci Trainer (dV-Trainer) (Mimic Technologies, Inc., Seattle, WA), has been recently adapted as a "back-pack" to be used with the Si console of the da Vinci robot, thus creating a third simulator, namely the da Vinci Surgical Skills Simulator (dVSSS) (Intuitive Surgical, Inc., Sunnyvale, CA). Although the construct validity of the dV-Trainer has been previously demonstrated, previous studies only included novices and experts, without a group of participants with intermediate level of robotic-assisted laparoscopic surgical expertise.⁹⁻¹¹ The advantage of the dVSS over the dV-Trainer is that it uses the same platform in the operating theatre, namely the Si console of the da Vinci robot. This means the trainee acquires skills using the same foot and finger clutch controls, as well as the endowrist manipulations that are used in the operating theatre. Although the dVSSS have been previously validated, these studies have shortcomings. Thus, the aim of this present pilot study was to confirm face, content and construct validity of the dVSSS using only 9 tasks with 3 levels of participants (novice, intermediate and expert) from 3 specialties of general surgery, gynecology and urology at a Canadian robotic-assisted laparoscopic surgery training centre.

Methods

After obtaining Institutional Research Ethics Board approval (Protocol 12-049), 48 participants were recruited across the specialties of urology, gynecology and general surgery between March 19 and April 17, 2012 (Table 1). This included attending surgeons, fellows (Post-graduate years [PGY] >5), residents (PGY 1-5) and medical students. Since previous studies on the learning curve for robotic-assisted laparoscopic radical prostatectomy (RALP) have shown that outcomes improve after 70 cases, experts were defined as those who had performed more than 75 robotic cases independently.¹² Therefore, participants who had performed between 1 and 74 robotic cases were classified in the intermediate group, while participants without any previous robotic experience were included in the novice group (Table 1).

The dVSSS using Mimic software (Mimic Technologies, Inc., Seattle, WA) has been previously described.¹³⁻¹⁶ Each

participant was provided with a manual to read. Novice group participants were provided one hands-on session for 10 minutes to teach them how to use the simulator's clutches and finger controls. There was also a short video provided by the simulator that preceded each task, explaining the objectives and the difficulty of each task. Each participant completed 9 tasks (Peg board level 2, match board level 2, needle targeting, ring and rail level 2, dots and needles level 1, suture sponge level 2, energy dissection level 1, ring walk level 3, and tubes) (Fig. 1). These were chosen by a focus group of experts representing various technical skills required for robotic surgery, including endowrist manipulation, camera control, clutching, dissection, energy control, fourth arm control, needle control, as well as basic and advanced needle driving. For each participant, the Mimic simulator software calculated a total score for each task, ranging from 0 (worst) to 100 (best).

The score is based on a set of predetermined metrics including the following: time to complete exercise, economy of motion, instrument collision, excessive instrument force, master work space range, drops, instruments out of view, missed targets, misplaced energy, broken vessels and blood loss volume. Not all metrics were measured for all tasks. The first attempt score was used in the present study to avoid effects of practice. A detailed score report was obtained for each participant after the first attempt of the task.

Immediately after completing the 9 tasks, each participant was asked to complete a questionnaire (Appendix 1). This was used to evaluate face validity (how realistic is the simulator?) and content validity (how useful is the simulator

Table 1. Participant demographics								
Group	Novice (n=30)	Intermediate (n=12)	Expert (n=6)					
Position:								
 Attending 	1	4	6					
Fellow	1	3	-					
 Resident 	13	5	-					
 Medical student 	13	-	-					
 Research assistant 	2	-	-					
Specialty:								
 Urology 	9	7	3					
 Gynecology 	6	3	3					
 General surgery 	2	2	-					
Median age (range) (years)	27.5 (2–56)	34.5 (27–48)	51 (37–53)					
Median experience (range) (years)								
 Surgical 	0 (0–25)	6 (3–17)	19 (9–28)					
 Laparoscopic 	0 (0–7)	4 (0–12)	11.5 (8–18)					
Robotic	0	2 (0–4)	4 (2–5)					
Median robotic cases performed (range)	0	9 (20–45)	250 (75–390)					



Fig. 1. The 9 da Vinci Surgical Skills Simulator (dVRSS) exercises used were as follows: (A) Pegboard level 2: to assess camera and endowrist manipulations when transferring rings from a pegboard onto 2 pegs on the floor. (B) Match board level 2: to assess clutching and endowrist manipulation when picking up objects and placing them into their corresponding places. The objective was to avoid excessive instrument strain. (C) Needle targeting: to assess endowrist manipulation and needle control when inserting needles through a pair of matching colored targets. (D) Ring and rail level 2: to assess camera control and clutching as wells endowrist manipulation when picking up 3 colored rings and guiding them along their matching colored railings. (E) Dots and needles level 1: to assess endowrist manipulation and needle control as well as needle driving technique when inserting a needle through several pairs of targets that have various spatial positions.
(F) Suture sponge level 2: to assess endowrist manipulation, camera control, needle control as well as needle driving technique of a sponge. (G) Energy and dissection level 1: to assess endowrist manipulation and dissection as well as energy control when cauterizing and cutting small branching blood vessels. (H) Ring walk level 3: to assess endowrist manipulation, camera control, clutching and fourth arm control when guiding a ring along a curved vessel and retracting obstacles using the fourth arm. (I) Tubes: to assess endowrist manipulation, needle control and needle driving technique when performing a tubular anastomosis. 2013 Intuitive Surgical, Inc.

as a training tool?). Overall percentile scores for each of the 9 tasks were compared among the 3 groups to determine construct validity.

Statistical analysis

Pairwise comparisons (novice vs. intermediate; intermediate vs. expert; novice vs. expert) of scores on the 9 tasks and 11 metrics were conducted using 2-sided Wilcoxon test. Using Bonferroni adjustment for multiple comparisons, a pvalue less than 1.67% (0.05/3) was considered as evidence against the null hypothesis (no difference between the two groups under comparison).

Results

There were a total of 48 participants: 30 (62.5%) were novices, 6 (12.5%) were experts (all 6 experts were attending surgeons) and 12 (25%) were intermediates (Table 1). There were 19 participants form urology, 12 from gynecology (including 2 gynecology research assistants), 4 from general surgery and 13 medical students. The experts had a median of 11.5 years of laparoscopic experience, including 4 years of robotic-assisted laparoscopic experience. The median number of robotic-assisted laparoscopic cases was 9 and 250 for the intermediate and expert groups, respectively (Table 1).

In terms of face validity, all participants, including the expert group, highly rated (7-8 out of 10 on a Likert scale) the dVSSS for overall realistic representation, 3-dimension-

al graphics, instrument handling and precision (Table 2). In terms of content validity, all participants, including the expert group, agreed that the dVSSS was useful in assessing and training trainees in robotic skills (scores of 8-10 out of 10). Furthermore, participants, including the expert group, agreed that the dVSSS should be part of a residency training program with robotic training (median score of 10 out of

Table 2: Face	and content	validity	scores	(medians	and
ranges)					

Face validity	All Groups scores	Expert scores
Realistic	8 (4–10)	7 (6–8)
 3-dimentional graphics 	8 (4–10)	7 (6–8)
 Instrument handling and ergonomics 	8 (6–10)	8 (6–9)
 Precision of the virtual reality platform 	8 (5–10)	7.5 (7–9)
Content validity		
 Useful for training residents 	9 (6–10)	8.5 (6–9)
 Useful for training fellows 	9 (5–10)	8 (6–9)
 All novices to robotic surgery should have training on the da Vinci Skills Simulator prior to performing robotic surgery on patients 	10 (6–10)	10 (8–10)
 Useful in assessing the resident's skill to preform robotic surgery 	8 (3–10)	8 (5–9)
 Useful in assessing the progression of the trainee's robotic skills 	8 (2–10)	8 (5–10)
 The da Vinci Skills Simulator should be included in all residency programs with robotic training 	10 (4–10)	10 (4–10)

Likert scale: 1 indicating definitely disagree and 10 indicating definitely agree.

10) (Table 2). In addition, the expert group recommended 10 (4-30) hours of training for senior residents and 8 (5-20) hours of training for fellows on the dVSSS.

For construct validity, experts outperformed novices in all 9 tasks (p < 0.05). Intermediates outperformed novices in 7 of 9 tasks (p < 0.05); there were no significant differences in the energy dissection level 1 and ring walk level 3 tasks. Finally, experts scored significantly better than intermediates in only 3 of 9 tasks (match board level 2, dots and needles and energy dissection level 1) (p < 0.05) (Table 3). When the different metrics were compared among the 3 groups, the novice group was significantly different from the intermediate and expert groups in time to complete exercise, economy of motion, instrument collision, excessive instrument force, and total task score (Table 4). Master workspace range was the only metric with significant differences between expert and intermediate groups and between expert and novice groups (Table 4). However, the scores for this metric between novice and intermediate groups were similar.

Discussion

.....

Currently, the American College of Surgeon requires candidates to pass the FLS curriculum to ensure they have acquired the basic laparoscopic skills during their training. However, there are no such standardized curricula in assessing robotic surgical skill; with the ongoing growth of robotic surgery, we are likely to see a curricula in the near future. Currently, there are 3 robotic simulators commercially available (RoSS, dV-Trainer and dVSSS). The aim of our study was to confirm the validation of the dVSSS using only 9 tasks with 3 levels of participants (novice, intermediate and expert) from 3 specialties of general surgery, gynecology and urology at a Canadian robotic-assisted laparoscopic surgery training centre.

Out of the 48 study participants, 19 (40%) were from urology, 12 (25%) were from gynecology, 4 (8%) were from general surgery, and 13 (27%) were medical students. Previous studies had an overwhelming representation from urology.¹³⁻¹⁵ The only other dVSSS validation study that included general surgeons is the one by Finnegan and colleagues.¹⁶ Therefore, the current study included a participant cohort that is more representative of the current robotic surgeons, including urologists, gynecologists and general surgeons. This is fitting given the recent expansion of robotic surgery within the fields of gynecology and general surgery. In the 2 previous studies the novice group exclusively consisted of medical students, in our study, the novice group includes residents, fellows and an attending surgeons with no robotic experience in addition to medical students.¹³⁻¹⁴ Therefore, the novice group in the present study may have better represented surgical trainees who are novice robotic surgeons.

Various tasks and repetitions have been used. For exam-

ple, Liss and colleagues used 3 tasks (peg board 1, peg board 2 and tubes) and Kelly and colleagues used 5 tasks (camera targeting 1, energy switching 1, threading rings 1, dots and needles 1 and ring and rail 1).¹⁴⁻¹⁵ On the other hand, Hung and colleagues used 10 tasks with 3 repetitions, while Finnegan and colleagues used 24 tasks.^{13,16} While more data are gathered by increasing the number of tasks and repetitions, these could create 2 issues. First, in their assessment study, Finnegan and colleagues found that fatigue and boredom occurred in cases in which it took up to 4 hours to perform the assessment.¹⁶ The second issue is the inadvertent effect of training while assessing the skill level of participants over multiple repetitions of the same task. Indeed, when Teishima and colleagues compared certified expert laparoscopic urologists with novice laparoscopic urologists, significant differences in their performance on the suture sponge task of the dV-Trainer were only seen at the second and third trials, while there were no significant differences between the two groups at the first and fourth attempts.¹⁷ This has several implications. First, even expert laparoscopic surgeons may need to become familiar with the robotic platform by performing a specific task. Second, some tasks may require few repetitions to achieve proficiency. Therefore, for the present study, the number of tasks were limited to 9 to avoid unnecessary fatigue and prolonged assessment period. Furthermore, the first attempt was used to avoid the effects of training.

In terms of face and content validities, all participants, including the expert group, highly rated the dVSSS as a realistic simulator useful for training residents in robotic skills (7-10 on a Likert scale of 1-10) (Table 2). This is similar to the findings of other studies.¹³⁻¹⁵ However, similar to other studies, the face and content validities of dVSSS were not compared to those of other robotic simulators, since the dVSSS was the only robotic simulator available at the study centre. While the advantage of dVSSS includes using the same robotic platform (console) used in the operating theatre, the same could also be a disadvantage since the console/simulator would only be available outside of operating theatre working hours. Therefore, the other simulators, such as the dV-Trainer and the RoSS, would be available during working hours for training. For the present study, the dVSSS was used only for assessment and training of robotic skills and thus it was not available for clinical cases.

The present study confirms the construct validity of the dVSSS in that experts outperformed novices in all 9 tasks (p < 0.05). Intermediates outperformed novices in 7 of 9 tasks (p < 0.05); there were no significant differences in the energy dissection level 1 and ring walk level 3 tasks. Finally, experts scored significantly better than intermediates in only 3 of 9 tasks (match board level 2, dots and needles and energy dissection level 1) (p < 0.05) (Table 3). Similarly, when metrics were compared among the 3 groups, there

Table 5: All 9 task scores stratmed by group (median and range)									
	Group								
Exercise	Novice n=30	Intermediate n=12	Expert n=6	Intermediate vs. novice (p value)	Expert vs. intermediate (<i>p</i> value)	Expert vs. novice (<i>p</i> value)			
Peg board (Level 2)	79 (31–98)	91.5 (76–99)	91.5 (85–99)	0.01	0.71	0.01			
Match board (Level 2)	64 (27–84)	75.5 (40–81)	85.5 (76–92)	0.03	0.01	<0.01			
Needle targeting	71.5 (6–90)	94 (71–100)	94 (86–100)	<0.01	0.64	<0.01			
Ring and rail (Level 2)	40 (8–75)	63 (32–91)	71 (53–87)	0.01	0.34	<0.01			
Dots and needles (Level 1)	57 (5–87)	80 (28–97)	93 (74–100)	0.01	0.05	<0.01			
Suture sponge (Level 2)	58 (21–83)	77 (53–97)	86.5 (72–92)	<0.01	0.57	<0.01			
Energy dissection (Level 1)	76 (36–97)	84 (61–97)	96 (87–99)	0.11	0.03	<0.01			
Ring walk (Level 3)	37.5 (0–78)	53 (6–82)	65.5 (39–88)	0.15	0.19	0.02			
Tubes	50 (28–79)	67 (42–99)	78 (53–83)	0.01	0.19	<0.01			

Table 3: All 9 task scores stratified by group (median and range)

were significant results between novice and intermediate groups and between novice and expert groups (Table 4). This is similar to the other two dVSSS validation studies where there were no significant differences between expert robotic surgeons and intermediates (senior residents).^{14,15} Finnegan and colleagues showed significant differences between the expert and intermediate groups in only 6 out of 24 exercises tested (dots and needles level 2, energy dissection level 1, energy switching, suture sponge level 2, thread the rings, and tubes).¹⁶ The only dVSSS validation study to show significant differences between intermediates and experts was the one by Hung and colleagues.¹³ However, given that study participants performed 3 repetitions for each of the 10 tasks, this may have increased the sample size of the collected data despite introducing a confounding effect of practice.

The lack of significant differences between the intermediate and expert groups on most of the exercises (6 out of 9) is a limitation of the present study and of the dVSSS. Perhaps if the intermediates and experts were allowed to practice prior to collecting data, their scores could have been significantly different in more tasks. In addition, 3 trials rather than only 1 trial would have resulted in narrower standard deviation and perhaps more significant difference between the intermediate and expert groups. In the present study, previous independent performance of 75 robotic cases was arbitrarily chosen as the cut-off between intermediate and expert group. Finnegan and colleagues showed that after 90 robotic cases, there was a plateau of the mean overall score on the dVSSS.¹⁶ Other studies have used cut-offs of 100 or 150 cases to define experts. Furthermore, in the present study, the maximum number of robotic cases performed was 390, whereas in other studies, the expert group contained participants with 800 to 1500 robotic cases. Another limitation of the present study is that the expert group was small

Table 4. Differences among groups in metrics measured								
			Grou	р				
Matrix	Novice	Intermediate	Expert	Novice vs. intermediate (p value)	Expert vs. intermediate (<i>p</i> value)	Expert vs. novice (p value)		
Time to complete exercise	2801 (1400–5823)	1795 (1258–3777)	1334 (1122–2327)	<0.01	0.17	<0.01		
Economy of motion	3357 (2352–9851)	2580 (1987–3546)	2178 (2160–2959)	<0.01	0.53	<0.01		
Instrument collisions	56 (23–180)	25 (12–80)	25 (10–37)	<0.01	0.56	<0.01		
Excessive instrument force	77 (77–554)	7 (1–164)	3 (0–70)	<0.01	0.86	0.03		
Master workspace range	87 (78–118)	88 (76–101)	74 (72–91)	0.14	0.04	<0.01		
Drop	1 (0–23)	1 (0–3)	0 (0–1)	0.25	0.63	0.16		
Instruments out of view	24 (1–335)	9 (2–107)	5 (0–71)	0.11	0.90	0.19		
Missed targets	42 (17–193)	24 (11–110)	15 (13–44)	0.03	0.41	0.01		
Misapplied energy time	13 (3–73)	10 (4–37)	8 (3–22)	0.37	0.39	0.11		
Broken vessels	0 (0–2)	0 (0–0)	0 (0–0)	0.22	1.00	0.34		
Blood loss volume	27 (0–228)	22 (0–256)	0 (0–11)	0.65	0.19	0.08		
Total task score	520 (178–686)	671 (460–799)	766 (642–823)	<0.01	0.18	<0.01		

(only 6 participants). It is expected that as robotic surgical experience advances, larger number of expert participants with larger robotic case logs would be included in future studies. Furthermore, the future generation of robotic simulators needs to incorporate more complex robotic tasks, such as robotic surgeries, with anatomical variations to better distinguish experts from intermediates.

Conclusion

Notwithstanding the limitations of our present pilot study, it confirms the face, content and construct validities of the dVSSS across the 3 surgical disciplines of urology, gynecology and general surgery. Larger sample size and more complex tasks are needed to further differentiate intermediates from experts.

Competing interests: This study was supported by an unrestricted educational grant from Abbott Pharmaceuticals and Paladin Labs Inc., Canada. Dr. Xianming Tan of the Biostatistics Core Facility, McGill University Health Centre Research Institute was commissioned to perform the statistical analysis.

This paper has been peer-reviewed.

References

- Healy GB. The college should be instrumental in adapting simulators to education. Bull Am Coll Surg 2002;87:10-1.
- Fried GM. FLS assessment of competency using simulated laparoscopic tasks. J Gastrointestsurg 2008;12:210-2. http://dx.doi.org/10.1007/s11605-007-0355-0
- Dauster B, Steinberg AP, Vassiliou MC, et al. Validity of the MISTELS simulator for laparoscopy training in urology. J Endourol 2005;19:541-5. http://dx.doi.org/10.1089/end.2005.19.541

- Sweet RM, Beach R, Sainfort F, et al. Introduction and validation of the American Urological Association Basic Laparoscopic Urologic Surgery skills curriculum. J Endourol 2012;26:190-6. http://dx.doi.org/10.1089/end.2011.0414
- Stefanidis D, Hope WW, Scott DJ. Robotic suturing on the FLS model possesses construct validity, is less physically demanding, and is favored by more surgeons compared with laparoscopy. Surg Endosc 2011;25:2141-6. http://dx.doi.org/10.1007/s00464-010-1512-1
- Medical Advisory Secretariat. Robotic-assisted minimally invasive surgery for gynecologic and urologic oncology: an evidence-based analysis. Ont Health Technol Assess Ser 2010;10:1-118. http://www.health. gov.on.ca/english/providers/program/mas/tech/reviews/pdf/rev_robotic_surgery_20101220.pdf. Accessed June 26,2013.
- Seixas-Mikelus SA, Kesavadas T, Srimathveeravalli G, et al. Face validation of a novel robotic surgical simulator. Urology 2010;76:357-60. http://dx.doi.org/10.1016/j.urology.2009.11.069
- Seixas-Mikelus SA, Stegemann AP, Kesavadas T, et al. Content validation of a novel robotic surgical simulator. BJU Int 2011;107:1130-5. http://dx.doi.org/10.1111/j.1464-410X.2010.09694.x
- Lendvay TS, Casale P, Sweet R, et al. VR robotic surgery: randomized blinded study of the dV-Trainer robotic simulator. Stud Health Technol Inform 2008;132:242-4.
- Kenney PA, Wszolek MF, Gould JJ, et al. Face, content, and construct validity of dV-trainer, a novel virtual reality simulator for robotic surgery. *Urology* 2009;73:1288-92. http://dx.doi.org/10.1016/j.urology.2008.12.044
- Sethi AS, Peine WJ, Mohammadi Y, et al. Validation of a novel virtual reality robotic simulator. J Endourol 2009;23:503-8. http://dx.doi.org/10.1089/end.2008.0250
- Raman JD, Dong S, Levinson A, et al. Robotic radical prostatectomy: operative technique, outcomes, and learning curve. JSLS 2007;11:1-7.
- Hung AJ, Zehnder P, Patil MB, et al. Face, content and construct validity of a novel robotic surgery simulator. J Urol 2011;186:1019-24. http://dx.doi.org/10.1016/j.juro.2011.04.064
- Kelly DC, Margules AC, Kundavaram CR, et al. Face, content, and construct validation of the Da Vinci Skills Simulator. Urology 2012;79:1068-72. http://dx.doi.org/10.1016/j.urology.2012.01.028
- Liss MA, Abdelshehid C, Quach S, et al. Validation, correlation, and comparison of the da vinci trainer and the da vinci surgical skills simulator using the mimi software for urologic robotic surgical education. *J Endourol* 2012;26:1629-34. http://dx.doi.org/10.1089/end.2012.0328
- Finnegan KT, Meraney AM, Staff I, et al. da Vinci Skills Simulator construct validation study: correlation of prior robotic experience with overall score and time score simulator performance. *Urology* 2012;80:330-5. http://dx.doi.org/10.1016/j.urology.2012.02.059
- Teishima J, Hattori M, Inoue S, et al. Impact of laparoscopic experience on the proficiency gain of urologic surgeons in robot-assisted surgery. J Endourol 2012;26:1635-8. http://dx.doi.org/10.1089/end.2012.0308

Correspondence: Dr. Maurice Anidjar, Sir Mortimer B. Davis Jewish General Hospital, 3755 Cote Ste-Catherine, Room E-945, Montreal, QC H3T 1E2;fax: 514-340-7559; maurice.anidjar@mcgill.ca

Appendix 1. Da Vinci Surgical Ski	Il Simulator Study Questionna	aire	
	[1	
Name		Age	
Demonstration		J	
Department]	
Position	Attending		
	Fellow		
	Resident —> Level R		7
	Medical student		
	Nurse		
	Other		
Years in practice	Surgical]
	Laparoscopic		
	Robotic		
			-
Previous robotic experiences			_
No. of cases performed at the console	e as primary surgeon		
	No. of cases in the last year		
			7
No. of cases performed at the console	e as assistant		
			1
Previous experience with	YES	NO	
		<u> </u>	
Previous experience with	\/ F O]
Da Vinci Skill Simulator	YES	NO	
			7
Previous experience with other	YES	NO	
robotic simulators			-
	Specify		

Appendix 1.	Da Vinci	i Surgical	Skill Sim	ulator Stud	ly Questio	nnaire (co	ntinued)			
From a scale	of 1 to 10), answer t	he following	g question.						
(If the scale of	doesn't ap	ply to you	, mark on "	not compete	ent to answ	er this ques	tion.")			
Is the skill si	mulator re	alistic?								Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disagree	e								Definitely agree	question
The 3D grap	nics are re	alistic.								Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disagree	e								Definitely agree	question
The instrum	nt handli	ng and are	ionomios ar	o roalistio						•• •
1	2 2	ng anu erg ຈ		E Teansile.	6	7	Q	Q	10	Not competent
I Definitely disagra	<u>د</u>	5	4	5	0	,	0	5	Definitely agree	question
Demittery disagre									Deminicity agree	4
The level of	orecision o	of the virtu	al reality pl	atform in th	e Skill Simi	ulator was r	ealistic.			Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disagree	e								Definitely agree	question
The Skill Sim	nulator is ι	useful for t	raining resi	dents in rot	ootic surger	y.				Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disagree	e								Definitely agree	question
The Skill Sim	nulator is ເ	useful for t	raining fello	ows in robo	tic surgery.	_	-			Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disagree	e								Definitely agree	question
All novices to	o robotic s	suraerv sh	ould have tr	aining on t	he Skill Sim	ulator prior	to perform	ina robo	tic surgerv	
on patients.										Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disagre	e								Definitely agree	question
The Skill Sim	nulator is u	useful in a	ssessing the	e resident's	skill to perf	orm robotic	surgery.			Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disagree	e								Definitely agree	question

Annondiv	1 Do Vinc	si Suraical	Skill Simi	ulator Stur	ly Questia	nnaire (co	ntinued)			
Аррениіх		Jurgica	SKIII SIIII		ly Questio		intinueu/			
The Skill S	imulator is	useful in as	ssessing the	e progressio	n of the tra	inee's robo	tic skills.			Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disa	gree								Definitely agree	question
The Skill S	imulator sh	ould be inc	luded in all	residency p	orograms w	ith robotic	training.			Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disa	gree								Definitely agree	question
As a reside	ent or fellov	v, training c	on the Skill	Simulator w	vill improve	your time of	on the cons	ole.		Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disa	igree								Definitely agree	question
Previous la	aparoscopic	experience	e helps in o	btaining ski	ils on the ro	botic simul	ator.	0	10	Not competent
l Deficient allo	2	3	4	5	6	/	8	9		to answer this
Definitely disa	igree								Definitely agree	question
Experience	e with video	games (e.g	g., PS, X-Bo	ox, etc) help	s improve t	he perform	ance with th	ne robotic	:	Not competent
simulator.	2	2	4	F	c	7	0	0	10	to answer this
I Dofinitaly diag	Z	3	4	5	0	/	ð	9	IU Definitely agree	question
Demnitely disa	igree								Demnitely agree	
The work of	on the Skill	Simulator h	nelps you u	nderstand r	obotic surg	ery.				Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disa	igree								Definitely agree	question
The work of	on the Skill	Simulator a	attracts you	to do robot	ic surgerv.					Not competent
1	2	3	4	5	6	7	8	9	10	to answer this
Definitely disa	igree								Definitely agree	question

Appendix 1. Da Vinci Surgical Skill Simulator Study Questionnaire (continued)

In your opinion, how many total hours should be spent on the Skill Simulator for the purpose of training prior to sitting on the surgical console?

Junior residents Senior residents

Fellows

What would you improve in the Skill Simulator?