

**Outcomes of the artificial urinary sphincter among men with prior pelvic radiation: Device survival and impact of time since radiation**Heather Rotz<sup>1</sup>, Blayne Welk<sup>1,2</sup><sup>1</sup>Department of Surgery, Western University, London, ON, Canada; <sup>2</sup>Department of Epidemiology and Biostatistics, Western University, London, ON, Canada**Cite as:** Rotz H, Welk B. Outcomes of the artificial urinary sphincter among men with prior pelvic radiation: Device survival and impact of time since radiation. *Can Urol Assoc J* 2025 January 14; Epub ahead of print. <http://dx.doi.org/10.5489/cuaj.9012>

Published online January 14, 2025

**Corresponding author:** Dr. Blayne Welk, Department of Epidemiology and Biostatistics, Western University, London, ON, Canada; [bkwelk@gmail.com](mailto:bkwelk@gmail.com)

\*\*\*

**INTRODUCTION**

The artificial urinary sphincter (AUS) improves continence and quality of life for men with post-prostatectomy incontinence.<sup>1</sup> Many of these prostate cancer patients also receive radiation therapy as primary treatment or salvage therapy. The exact impact of radiation on outcomes of the AUS is still debatable, but in general studies suggest that radiation results in higher risk of erosion and reoperation.<sup>2-4</sup>

Radiation disrupts the basement membrane of blood vessels, leading to microangiopathy, neovascularization and subsequently causes fibroblast proliferation and scar formation. These effects typically manifest in a delayed and cumulative fashion, with reports of complications such as fistulas, urethral strictures, and hemorrhagic cystitis presenting decades after the completion of radiation therapy.<sup>5</sup> Our objective was to describe our surgical outcomes in AUS patients with prior pelvic radiation, and determine if time since radiation is significantly related to reoperative device complications; our hypothesis was that increased time from radiation would be associated with increased AUS complications.

**KEY MESSAGES**

- Patients undergoing artificial urinary sphincter (AUS) implantation often have received radiation; we did not find that the time since radiation impacts AUS complications that require reoperation.
- Despite the challenges with operating in radiated patients, the five-year AUS device survival was high.
- A larger AUS cuff size is significantly correlated with a lower risk of AUS complications that require reoperation.

**METHODS**

This study is a retrospective case series. We used billing records to identify consecutive adult males who underwent a first time AUS between 2012 to 2023 with a single surgeon. A two-incision technique was used, and a perineal incision was used for proximal bulbar cuff placement. A transcorporeal approach, or grafts/bulbospongiosus wraps were not used in any of these first time AUS cases. We only included patients who had received pelvic radiation prior to AUS. Ethics approval was granted by Western University REB (120719)

We used a standardized data extraction sheet to collect data from the paper and electronic patient charts, and a sample of data was double extracted to ensure accuracy. The primary outcome was an AUS complication requiring repeat surgical intervention. This included an erosion and/or device infection, a revision surgery for recurrent urinary incontinence or other device problems. Our secondary outcome was device reoperation due to infection/erosion (which may be more specific to a radiation related mechanism). The primary exposure was time, in months, from completion of radiation to AUS insertion. We performed exploratory analysis on secondary exposure variables, such as cuff size, learning curve, and American Society of Anesthesiologists (ASA) score. For all analyses, cox regression analysis was used and patients were censored at last follow-up or death. Due to our limited sample size and low event rate, we could not create strong multivariable models, therefore we only reported univariable cox regression models. SPSS version 26 was used for statistical analysis, and a  $p < 0.05$  was considered statistically significant.

## RESULTS

We identified a total of 99 men (out of 200 total first time AUS placements) who met inclusion criteria. Relevant baselines, intraoperative details and outcomes are shown in table 1. The most used cuff size was 4 cm (43%). Radiation dosage and fractions were available for 52/99 patients, and almost all men received 66 gray (Gy) in 33 fractions. The median time between radiation and AUS implantation was 90 (IQR 41-129) months.

After AUS placement, 34 men (34%) had a Clavien  $\leq 3$  complication within 30 days (such as urinary retention, urinary tract infection, or gross hematuria), and 92% of men had social continence. After a median follow-up of 31 months, 15 men (15%) had a repeat surgical intervention on their AUS. The majority (12/15) were for device erosion/infection. The remaining 3 were for mechanical failure (1) and recurrent incontinence (2). The 5-year overall device survival was 87%.

Time in months from radiation was not significantly associated with all-cause device failure (HR 1.00, 95% CI 0.99-1.01,  $p=0.63$ ), or with our secondary outcome specific to infection/erosion (HR 1.0, 95%CI 0.99-1.01,  $p=0.44$ ). We explored other exposure variables and found that larger cuff size was significantly associated with lower risk of all-cause AUS reoperation (HR 0.27, 0.05-0.91,  $p=0.03$ ) and a higher ASA class was significantly associated with higher risk of all-cause AUS reoperation (HR 5.28, 1.42-19.70,  $p=0.01$ ). Of the 15 people with reoperation, 13 (87%) had a 4.0 or 3.5cm cuff, and 12 (80%) were ASA 3 or 4.

## DISCUSSION

While the increased risk of complications and need for surgical revision following AUS implantation in previously radiated patients has been described, predicting this risk remains elusive. The delayed complications of radiation therapy can be severe, and the resulting fibrosis, tissue atrophy, and impaired wound healing increases the potential risk of surgery. The onset of these complications typically manifests between six months to several years post-treatment, particularly with doses  $\geq 65$  Gy, which is common in prostate cancer.<sup>6</sup> The purpose of this study was to explore if there was an ideal window after radiation during which to place an AUS; to our knowledge has not been investigated previously. We did not find that the length of time since radiation impacts the risk of an AUS complication.

In general, radiation has been shown to accelerate time to cuff erosion. In a multicentre series, mean overall time from AUS implant to erosion was shortened to one year with radiation compared with three years in patients with no prior radiation.<sup>7</sup> The risk of AUS complication seems to increase with pelvic radiation, and may be more likely if the AUS is placed after radiation rather than prior to radiation.<sup>8</sup> This suggests that the ideal time to place an AUS may be prior to the radiation therapy, when tissue changes have not yet occurred, and once a patient has been radiated, according to our results, the timing of AUS placement does not impact significant device complications.

We found that larger AUS cuff size was protective against reoperation complications in radiated patients. This is similar to the results of a large retrospective study of 1020 patients that found that cuff size (but not history of radiation) was significantly related to device failure.<sup>9</sup> Other studies have demonstrated that a very small cuff sizes (3.5cm) in combination with radiation leads to an unacceptably high erosion risk of 85% (or 6.2 fold higher).<sup>10</sup> The larger cuff size likely represents a healthier urethra, which explains the association between smaller cuff size and urethra erosion/infection. We abandoned the 3.5cm cuffs early in practise, so we are unable to study this cuff size specifically. We also found that higher ASA score was significantly associated with AUS complications, which is likely driven by poor wound healing and impaired circulatory function that often accompanies patients with a higher ASA score.

The current study is limited by its retrospective nature. It is not possible to determine the impact of very early AUS placement after radiation, as it not our practise to place an AUS until after at least 6 months after radiation therapy is completed. The risk of surgical complications was low, and thus we had limited power to assess our primary outcome, and could not do multivariable modelling. This study represents data from a single surgeon at a single institution and findings may not be generalizable to other settings.

## CONCLUSIONS

Patients with prior radiation have a high device survival at 5 years. Time since radiation is not a significant predictor of device failure, however a larger cuff size and lower ASA score may be protective for repeat surgical intervention.

## REFERENCES

1. Van der Aa F, Drake MJ, Kasyan GR, et al. The artificial urinary sphincter after a quarter of a century: A critical systematic review of its use in male non-neurogenic incontinence. *Eur Urol* 2013;63:681-9. <https://doi.org/10.1016/j.eururo.2012.11.034>
2. Fuller TW, Ballon-Landa E, Gallo K, et al. Outcomes and risk factors of revision and replacement artificial urinary sphincter implantation in radiated and nonradiated cases. *J Urol* 2020;204:110-4. <https://doi.org/10.1097/JU.0000000000000749>
3. Bates AS, Martin RM, and Terry TR. Complications following artificial urinary sphincter placement after radical prostatectomy and radiotherapy: A meta-analysis. *BJU Int* 2015;116:623-33. <https://doi.org/10.1111/bju.13048>
4. Srivastava A, Joice GA, Patel HD, et al. Impact of adjuvant radiation on artificial urinary sphincter durability in postprostatectomy patients. *J Reconstr Urol* 2018;114:212-7. <https://doi.org/10.1016/j.urology.2017.12.029>
5. Cooper JS, Hanley ME, Hendriksen S, et al. Hyperbaric treatment of delayed radiation injury. *StatPearls* 2023 Jul.
6. Feldmeier JJ. Hyperbaric oxygen therapy and delayed radiation injuries (soft tissue and bony necrosis): 2012 update. *Undersea Hyperb Med* 2012;39:1121-39.
7. Kaufman MR, Milam DF, Johnsen NV, et al. Prior radiation therapy decreases time to idiopathic erosion of artificial urinary sphincter: A multi-institutional analysis. *J Urol* 2018;199:1037-41. <https://doi.org/10.1016/j.juro.2017.11.046>
8. DeLay KJ, Haney NM, Chang J, et al. Comparison of adjuvant radiation therapy before or after artificial urinary sphincter placement: A multi-institutional, retrospective analysis. *J Urol* 2018;113:160-5. <https://doi.org/10.1016/j.urology.2017.11.041>
9. Bentellis El-Akri M, Cornu J-N, et al. Prevalence and risk factors of artificial urinary sphincter revision in nonneurological male patients. *J Urol* 2021;206:1248-57. <https://doi.org/10.1097/JU.0000000000001954>
10. Simhan J, Morey AF, Singla N, et al. 3.5 cm artificial urinary sphincter cuff erosion occurs predominantly in irradiated patients. *J Urol* 2015;193:593-7. <https://doi.org/10.1016/j.juro.2014.07.115>

## FIGURES AND TABLES

<b>Table 1. Patient demographic details</b>		<b>Median (IQR) or n</b>
Age (years)		72 (69–76)
BMI (kg/m <sup>2</sup> )		29 (26–31)
ASA class		
1		2
2		29
3		63
4		5
Etiology of incontinence		
EBRT alone		1
EBRT + RRP		89
Brachy + TURP		6
RT + Other		3
Patients with prior non-AUS incontinence procedures		4
Time between radiation and AUS placement (months)		90 (41–129)
Patients with vesicourethral anastomotic stricture (defined as <12 Fr)		13
Patient reported pads per day prior to the AUS		5.0 (3.0–7.3)
Cuff size (cm)		
3.5		3
4		43
4.5		37
5		10
>5		6
30 postop day complications		
Clavien 1–2		22
Clavien 3		12
Clavien 4–5		0
Social continence after activation (defined as ≤1 pad per day)		92
Total followup time (median, IQR)		31 (12–60)

ASA: American Society of Anesthesiologists; AUS: artificial urinary sphincter; BMI: body mass index; EBRT: external beam radiation therapy; IQR: interquartile range; RRP: retropubic radical prostatectomy; RT: radiation therapy; TURP: transurethral resection of the prostate.

<b>Table 2. Univariate analysis of variables and their association with AUS reoperation</b>
---

	<b>HR (95% CI)</b>	<b>p</b>
Impact of time from radiation on any AUS reoperation	1.00 (0.99–1.01)	0.63
Impact of time from radiation on AUS reoperation for erosion/infection	1.00 (0.99–1.01)	0.44
Exploratory analyses of variables and their relationship to any AUS reoperation		
Age (per year increase)	1.02 (0.93–1.13)	0.64
BMI (per kg/m <sup>2</sup> increase)	0.92 (0.78–1.09)	0.33
Vesicourethral anastomotic stricture (relative to those without VUAS)	2.31 (0.68–7.88)	0.18
Cuff size (per cm increase)	0.27 (0.05–0.91)	0.03
ASA class (per unit increase)	5.28 (1.42–19.70)	0.01
Learning curve (per case number)	1.02 (0.99–1.06)	0.13

ASA: American Society of Anesthesiologists; AUS: artificial urinary sphincter; CI: confidence interval; HR: hazard ratio.

DRAFT